



THE ASSOCIATION OF ANAESTHETISTS
of Great Britain & Ireland

WILEY-BLACKWELL

AAGBI Core Topics in Anaesthesia 2012

Edited by Ian Johnston, William Harrop-Griffiths & Leslie Gemmell



Contents

List of Contributors

Foreword

CHAPTER 1: The Physics of Ultrasound

Sound energy

Generation of the ultrasound wave

Refinements in transducer production

The frequency, wavelength and speed of sound waves

Frequency

Wavelength

Speed

The image

Reduced and distorted images

The probe

Terminology of modes

Doppler shift

The safety of ultrasound

CHAPTER 2: Coronary Artery Stents: Management in Patients Undergoing Noncardiac Surgery

Percutaneous coronary interventions

Radial or femoral access

Antiplatelet and anticoagulant therapy

Coronary artery stenting and noncardiac surgery
Management of complications in patients with stents
Adjunct technology and future developments in coronary stenting
Aspirin and clopidogrel resistance
Conclusions

CHAPTER 3: Anaesthesia and Enhanced Recovery for Colorectal Surgery

Preoperative preparation
Admission
Intraoperative components of ERAS
Postoperative components of ERAS
Discharge
Audit and outcome measurement
Quality Innovation Productivity and Prevention agenda
Summary

CHAPTER 4: The Unanticipated Difficult Airway: The ‘Can’t Intubate, Can’t Ventilate’ Scenario

Why does this scenario occur?
Avoiding the CICV situation
Cricothyroidotomy
Wide bore or narrow bore cannula?
Complications of cricothyroidotomy
Training in cricothyroidotomy

CHAPTER 5: Analgesia for Abdominal Surgery

Epidural analgesia

Spinal anaesthesia

Rectus sheath block

Transversus abdominis plane block

Ilioinguinal and iliohypogastric blocks

Paravertebral block

Summary

CHAPTER 6: Analgesic Regimens for Children

Pain perception

Designing analgesic regimens

Analgesic drugs and techniques

Chronic pain

Postoperative nausea and vomiting (PONV)

CHAPTER 7: The Unstable Cervical Spine

The provision and maintenance of a reliable, patent airway is paramount

Maintenance of spinal cord perfusion is a priority

Airway techniques that minimally displace the alignment of the vertebrae should be used

Normal cervical spine anatomy

Definition of instability

Neurological consequences of instability

Causes of cervical instability

Diagnosis

Conservative management

Surgical management

Anaesthetic management

Airway management

Spinal cord perfusion

CHAPTER 8: Obstetric Haemorrhage

Epidemiology

Antepartum haemorrhage: causes

Postpartum haemorrhage: causes

Major haemorrhage: immediate management

Management in theatre

Options for haemostasis: avoidance of hysterectomy

Emergency anaesthesia

Elective caesarean section: anticipated major haemorrhage

Anaphylactoid syndrome of pregnancy

Cardiac arrest

CHAPTER 9: Anaesthesia for Patients Undergoing Hip Fracture Surgery

Demographics

Preoperative management

The timing of surgery

Intraoperative anaesthetic care

Postoperative care

CHAPTER 10: e-Learning Anaesthesia

[The rationale behind e-Learning Anaesthesia](#)
[The e-Learning Anaesthesia project](#)
[Summary](#)

[CHAPTER 11: Consent and the Explanation of Risk in Anaesthesia](#)

[Common law of consent](#)
[Competence](#)
[Patients who lack competence](#)
[Children](#)
[Teaching, learning and maintaining skills](#)
[Research](#)
[Data protection](#)
[Needlestick injuries](#)
[Problems with consent](#)

AAGBI Core Topics in Anaesthesia

AAGBI Core Topics in Anaesthesia

EDITED BY

Ian Johnston

Raigmore Hospital
Inverness, UK

William Harrop-Griffiths

Imperial College Healthcare NHS Trust
London, UK

Leslie Gemmell

Wrexham Maelor Hospital
Wrexham, UK



WILEY-BLACKWELL

A John Wiley & Sons, Ltd., Publication

This edition first published 2012 © The Association of
Anaesthetists of Great Britain and Ireland (AAGBI)

Wiley-Blackwell is an imprint of John Wiley & Sons, formed
by the merger of Wiley's global Scientific, Technical and
Medical business with Blackwell Publishing.

Registered Office

John Wiley & Sons, Ltd, The Atrium, Southern Gate,
Chichester, West Sussex, PO19 8SQ, UK

Editorial Offices

9600 Garsington Road, Oxford, OX4 2DQ, UK
The Atrium, Southern Gate, Chichester, West Sussex, PO19
8SQ, UK

350 Main Street, Malden, MA 02148-5020, USA

For details of our global editorial offices, for customer
services and for information about how to apply for
permission to reuse the copyright material in this book
please see our website at www.wiley.com/wiley-blackwell

The right of the author to be identified as the author of this
work has been asserted in accordance with the UK
Copyright, Designs and Patents Act 1988.

All rights reserved. No part of this publication may be
reproduced, stored in a retrieval system, or transmitted, in
any form or by any means, electronic, mechanical,
photocopying, recording or otherwise, except as permitted
by the UK Copyright, Designs and Patents Act 1988, without
the prior permission of the publisher.

Designations used by companies to distinguish their
products are often claimed as trademarks. All brand names
and product names used in this book are trade names,
service marks, trademarks or registered trademarks of their
respective owners. The publisher is not associated with any
product or vendor mentioned in this book. This publication is
designed to provide accurate and authoritative information

in regard to the subject matter covered. It is sold on the understanding that the publisher is not engaged in rendering professional services. If professional advice or other expert assistance is required, the services of a competent professional should be sought.

The contents of this work are intended to further general scientific research, understanding, and discussion only and are not intended and should not be relied upon as recommending or promoting a specific method, diagnosis, or treatment by physicians for any particular patient. The publisher and the author make no representations or warranties with respect to the accuracy or completeness of the contents of this work and specifically disclaim all warranties, including without limitation any implied warranties of fitness for a particular purpose. In view of ongoing research, equipment modifications, changes in governmental regulations, and the constant flow of information relating to the use of medicines, equipment, and devices, the reader is urged to review and evaluate the information provided in the package insert or instructions for each medicine, equipment, or device for, among other things, any changes in the instructions or indication of usage and for added warnings and precautions. Readers should consult with a specialist where appropriate. The fact that an organisation or Website is referred to in this work as a citation and/or a potential source of further information does not mean that the author or the publisher endorses the information the organisation or Website may provide or recommendations it may make. Further, readers should be aware that Internet Websites listed in this work may have changed or disappeared between when this work was written and when it is read. No warranty may be created or extended by any promotional statements for this work. Neither the publisher nor the author shall be liable for any damages arising herefrom.

Library of Congress Cataloging-in-Publication Data

AAGBI core topics / edited by Ian G. Johnston, William Harrop-Griffiths, Leslie Gemmell.

p. cm.

Association of Anaesthetists of Great Britain and Ireland
core topics
Core topics

Includes bibliographical references and index.

ISBN-13: 978-0-470-65862-8 (pbk. : alk. paper)

ISBN-10: 0-470-65862-2 (pbk. : alk. paper)

1. Anesthesia. 2. Surgery, Operative. 3. Operations,
Surgical. I. Johnston, Ian G.

II. Harrop-Griffiths, William. III. Gemmell, Leslie. IV.
Association of Anaesthetists of Great Britain
and Ireland. V. Title: Association of Anaesthetists of Great
Britain and Ireland core topics. VI. Title: Core topics.

[DNLM: 1. Anesthesia-methods. 2. Anesthesia-
contraindications. 3. Surgical Procedures,
Operative. WO 200]

RD81.A23 2012

617.9'6-dc23

2011024801

A catalogue record for this book is available from the British
Library.

List of Contributors

Graham Arthurs

Wrexham Maelor Hospital
Wrexham, UK

Nick Boyd

Derriford Hospital
Plymouth, UK

Ian Calder

The National Hospital for Neurology
and Neurosurgery
London, UK

Leslie Gemmell

Wrexham Maelor Hospital
Wrexham, UK

Alex Grice

Royal Devon and Exeter NHS
Foundation Trust
Exeter, UK

Richard Griffiths

Peterborough City Hospital
Peterborough, UK

Ed Hammond

Royal Devon and Exeter NHS
Foundation Trust
Exeter, UK

William Harrop-Griffiths

Imperial College Healthcare NHS Trust
London, UK

Ian Johnston

Raigmore Hospital
Inverness, UK

Michelle Leemans

The National Hospital for Neurology
and Neurosurgery
London, UK

Stephen Leslie

Raigmore Hospital
Inverness, UK

David Levy

Nottingham University Hospitals
NHS Trust
Queen's Medical Centre Campus
Nottingham, UK

Andrew McIndoe

University Hospitals Bristol NHS
Foundation Trust
Bristol, UK

Simon Marshall

Musgrove Park Hospital
Taunton, UK

Colin Moore

Royal Infirmary of Edinburgh
Edinburgh, UK

Christopher Newell

University Hospitals Bristol NHS
Foundation Trust
Bristol, UK

Carol Peden

Royal United Hospital
Bath, UK

Mansukh Popat

Oxford Radcliffe Hospital
NHS Trust
Oxford, UK

Stuart White

Brighton and Sussex University Healthcare
NHS Trust
Brighton, UK

Glyn Williams

Great Ormond Street Hospital for Children
NHS Trust
London, UK

Foreword

Iain Wilson, President of AAGBI

The Association of Anaesthetists of Great Britain and Ireland has worked tirelessly since 1932 to promote and advance patient safety by offering anaesthetists the educational materials they need to support safe and effective practice. Until recently, most of our educational output took the form of the journal *Anaesthesia*, Seminars at 21 Portland Place, London, and meetings such as WSM London and the Annual Congress. However, times change and the AAGBI is well placed to adapt to changing times and the changing needs of its members. The development of the appraisal process, the start of Revalidation and the introduction of the CPD Matrix have all changed the way that Consultant Anaesthetists and Specialty Doctors need to structure and record their learning and continuous professional development.

Two years ago, we started a highly successful series of Core Topics Meetings that are held in towns and cities throughout the UK and Ireland. Our new website will give us opportunities to develop and deliver online education, and the Tutorial of the Week developed in collaboration with the World Federation of Societies of Anaesthesiologists is already available in pdf format from www.aagbi.org. This Core Topics book is another way in which we are offering the very best of educational material free of charge to our members. We have asked acknowledged experts to write concise and informative articles on topics that really are core to the safe delivery of patient care: difficult airways, pain management, hip fractures and obstetric haemorrhage to mention but four.

We know that our members use different educational material in different ways, and we think that even in a world of smart phones, blogs and social networking, many anaesthetists still favour the portability, convenience – and legibility – of a book. We are therefore offering this Core Topics book free of charge to our members to gauge how well it is received. We will be asking you for your feedback and, if it is positive and encouraging, we will consider creating a regular series of publications such as this.

I hope you enjoy reading this AAGBI Core Topics in Anaesthesia book. If you have any comments to make about it, any suggestions for future topics or if you would like to offer to write an article for us, please email me at president@aagbi.org.

CHAPTER 1

The Physics of Ultrasound

Graham Arthurs

Maelor Hospital, Wrexham, UK

Key points

- Ultrasound is a high-frequency pressure energy wave transmitted longitudinally through the soft tissues of the body.
 - Advances in computer technology have made medical ultrasound possible by processing millions of signals every second.
 - Ultrasound makes it possible to examine most of the tissues of the body safely and easily.
 - The pressure, energy and heating effects of clinical ultrasound devices have not been shown to damage normal biological tissues.
 - The ultrasound wave must be reflected off a tissue interface at right angles. This means that a combination of good hand-eye coordination and correct positioning of the probe is the basis of a good image.
 - Images are presented as patterns on a greyscale monitor. Pattern recognition is therefore the basis of the interpretation of these images.
 - The B or brightness mode gives a greyscale image that is distorted because of a loss of reflected echoes by scatter and refraction.
 - Doppler shift is caused by a change in wavelength when fluid such as blood is moving towards or away from the ultrasound wave.
-

This chapter aims to give an introduction to the basic physics of ultrasound in order to allow the reader to understand how images are produced and hence how to obtain the best images when using ultrasound. Clinical ultrasound devices simultaneously produce and transmit multiple pressure waves, and receive and rapidly interpret the many attenuated, returning pressure wave signals. The pressure wave signals are converted into electrical signals. The production of an image in real time by a portable

scanner has only become possible with the development of the microprocessor chip. A modern ultrasound device has a lot of computing power with many microprocessors performing many billion operations per second, which makes it possible to build up complicated images in real time. In the near future, smaller processors will enable more precise images to be created on smaller, lighter and cheaper devices.

Sound energy

Sound is energy in the form of a pressure wave. This mechanical energy displaces molecules that press against adjacent molecules and pass the energy on in waves, while the first molecules to be compressed together return to their original configuration. Ultrasound waves travel in a longitudinal direction, rather like a piston pumping in and out, producing a series of high pressure, unidirectional waves in front of the piston. The effect can be visualised as the energy of the wind blowing through a cornfield. The corn bends over but returns to the upright as the wind blows through the field. The pressure recorded in ultrasound waves can be as high as 10 times atmospheric pressure but it is a transient change. A sound wave needs molecules to pass the wave on so it cannot travel through a vacuum, and medical ultrasound does not pass air so, for instance, the lung will appear black unless there is an effusion or consolidation within the lung. The principle behind obtaining information using ultrasound is to send a short burst of pressure waves into the tissue, wait for the initial wave to bounce back from a number of tissue interfaces, and then analyse the waves that return. This can be understood with reference to the early application of sound waves in Sonar. A single wave of sound is sent out from a boat into the water and the time taken for the echo to return is measured.

Knowing the speed of sound in water and the time taken for the sound to return, the distance travelled by the sound can be calculated; the point at which the sound is reflected is half this distance.

Generation of the ultrasound wave

In order to build up an image of the tissues, it is necessary to have many sound waves transmitted from a series of points fixed in relationship to each other. This is made possible with a line or array of transducers in the probe head. The probe is the most delicate part of the device and a typical array contains 128-256 transducer elements in a wide aperture device. There are matrix arrays now on the market with up to 2400 transducer elements. Each transducer has a dual function. It both creates and sends out a short burst of ultrasound pressure waves, and is then silent until it detects those waves returning. This is referred to as pulsed ultrasound. Small groups of transducers are activated in turn to emit at the same time followed by another group. In this way some part of the probe head is always emitting and some part is always receiving. In order to build up an image, it is necessary to emit and receive many signals in a short period of time. This is made possible by the use of materials that exhibit the piezoelectric effect. Pierre and Jacques Curie found that a crystal of quartz changes shape when an electrical charge or voltage is applied across it and, conversely, when pressure is applied to the crystal, it changes its shape and an electrical charge is created. This two-way process of an electrical charge causing a change in shape that produces a pressure wave, and pressure causing a change in shape that creates an electrical charge, is known as the piezoelectric effect. The

emitted ultrasound signal is produced by applying an electrical charge to the transducer and the returning pressure wave distorts the transducer to create an electrical charge that can then be processed. The sensitivity of the transducer has been increased by replacing quartz crystals with materials such as lead zirconate titanate (PZT) mixed with epoxy, and polyvinylidene difluoride (PVDF). The transducer is made by heating powdered PZT to above its Curie temperature (365°C for PZT), sometimes mixed with epoxy to make a composite ceramic. While the crystals are hot and mobile a high voltage current is applied across the transducer to polarise the crystals so that they line up to give a maximum response to changes when an external electrical charge or pressure is applied. The voltage is maintained until the mixture cools and solidifies so that the alignment of the crystals is preserved in the solid state.

Refinements in transducer production

Each transducer has a particular frequency at which it will convert electrical energy into sound waves and vice versa. This is the *resonance frequency*, which is mostly dependent on the thickness of the transducer element: the thinner the element, the higher the frequency. In practice, the emitted wave is not a single wavelength but a spectrum of frequencies. These are produced by forming multiple matching layers on the surface of the element. These layers also provide good sound transmission between the transducer and the soft tissues over a range of frequencies. Modern transducers are constructed of piezoelectric materials that have similar acoustic impedances to that of soft tissue so that sound is not lost by reflection at the soft tissue-transducer interface.

The frequency, wavelength and speed of sound waves

A sound wave has characteristics of frequency, wavelength and speed. The frequency and speed of the sound wave determine the type of substance through which it will travel.

Frequency

The frequency (f) is the number of high-pressure waves in one second. The time taken from one high pressure to the next high pressure is known as the *period* (T). Therefore, $T = 1/f$

Frequency is quoted in Hz where $1 \text{ Hz} = 1 \text{ cycle.s}^{-1}$. Low-frequency sound waves or *seismic* waves are transmitted through dense materials such as the earth. Sound waves that are below the audible range are called *infrasonic*. Sound that is heard has a frequency in the range 20 Hz to 20 kHz. This is called the *audible frequency range*. Audible sound waves are transmitted through air at a speed of 330 m.s^{-1} . Ultrasound has a frequency above the audible range at 20 kHz. Clinical ultrasound is in the range of 2–15 MHz. This is transmitted well through the soft tissues of the body. Air-filled cavities and solid material such as bone and metal needles will not transmit this frequency of sound, so the lung and intracranial structures cannot be examined, except in babies when ultrasound can be focused through the fontanelles.

Wavelength

The wavelength (λ) is the distance between two high-pressure peaks in the sound wave. The wavelength is the speed of the sound wave divided by the frequency of the high-pressure components, hence $\lambda = c/f$. The wavelength is important for *spatial resolution*. This is the ability to identify two objects when one is deeper than the other (longitudinal spatial resolution) or side-by-side (axial spatial resolution) but only separated by a short distance. In soft tissues, wavelengths of about 1 mm are used (at 2 MHz the wavelength is 0.77 mm). A higher frequency produces a shorter wavelength. The width of the wave is about the same as the wavelength, so the width of the sound wave becomes smaller as the frequency increases. If the object presents a surface area that is less than the width of the sound wave, such as a red blood cell, the wave will not be reflected at 180°.

Speed

The speed of the sound wave is determined by the nature of the substance or organ it passes through. The speed $c = \sqrt{B/\rho}$, where B is the stiffness or *bulk modulus* and ρ is the density of the tissue or organ. The optimum speeds of sound for the different soft tissues of the body are:

- liver 1580 m.s⁻¹
- muscle 1575 m.s⁻¹
- blood 1570 m.s⁻¹
- water 1480 m.s⁻¹
- fat 1430 m.s⁻¹

A convention exists in which manufacturers of medical ultrasound devices all use the same sound wave speed to allow images to be comparable. This speed is 1540 m.s⁻¹.

Therefore, clinical ultrasound will not pass through materials that require a much higher or lower speed for wave transmission:

- air 330 m.s^{-1}
- lead 1240 m.s^{-1}
- bone 2800 m.s^{-1}
- skull bone 4080 m.s^{-1}
- aluminum 6400 m.s^{-1}

Medical ultrasound waves are all reflected off the surface of normal bone, so the internal structure of bone cannot be examined. It is possible to diagnose a fracture by examining the surface image of the bone, which may become irregular at the fracture site, or the clear surface image may be lost, indicating that the sound waves are entering the haematoma produced by the fracture. Needles are metal but unlike bone a clear surface image will not be seen because the surface area is small and curved, so very little of the wave will be focused on the needle and, even if it is focused it, will then be scattered into the tissue (see *axial resolution* and *scatter* below).

The fundamental properties of medical ultrasound are therefore:

- speed = 1540 m.s^{-1}
- frequency (f) = 2-12 MHz
- wavelength (λ) = 0.1-0.8 mm

The image

The image is presented as a greyscale. By convention, each returning signal lightens the black of the screen, with strong echoes showing as white and no echo as black. To get the best image, as much energy as possible in the emitted waves needs to return to the detector.