

Journal of
Prosthodontics
on Complex Restorations



Edited by
Nadim Z. Baba
David L. Guichet

JOURNAL OF PROSTHODONTICS
ON COMPLEX RESTORATIONS

JOURNAL OF PROSTHODONTICS ON COMPLEX RESTORATIONS

Edited by

NADIM Z. BABA, DMD, MSD, FACP

DAVID L. GUICHET, DDS, FACP



AMERICAN COLLEGE OF
PROSTHODONTISTS
Your smile. Our specialty.®

WILEY Blackwell

Copyright © 2016 by American College of Prosthodontists. All rights reserved.

Published by John Wiley & Sons, Inc., Hoboken, New Jersey.

Published simultaneously in Canada.

No part of this publication may be reproduced, stored in a retrieval system, or transmitted in any form or by any means, electronic, mechanical, photocopying, recording, scanning, or otherwise, except as permitted under Section 107 or 108 of the 1976 United States Copyright Act, without either the prior written permission of the Publisher, or authorization through payment of the appropriate per-copy fee to the Copyright Clearance Center, Inc., 222 Rosewood Drive, Danvers, MA 01923, (978) 750-8400, fax (978) 750-4470, or on the web at www.copyright.com. Requests to the Publisher for permission should be addressed to the Permissions Department, John Wiley & Sons, Inc., 111 River Street, Hoboken, NJ 07030, (201) 748-6011, fax (201) 748-6008, or online at <http://www.wiley.com/go/permission>.

The contents of this work are intended to further general scientific research, understanding, and discussion only and are not intended and should not be relied upon as recommending or promoting a specific method, diagnosis, or treatment by health science practitioners for any particular patient. The publisher and the author make no representations or warranties with respect to the accuracy or completeness of the contents of this work and specifically disclaim all warranties, including without limitation any implied warranties of fitness for a particular purpose. In view of ongoing research, equipment modifications, changes in governmental regulations, and the constant flow of information relating to the use of medicines, equipment, and devices, the reader is urged to review and evaluate the information provided in the package insert or instructions for each medicine, equipment, or device for, among other things, any changes in the instructions or indication of usage and for added warnings and precautions. Readers should consult with a specialist where appropriate. The fact that an organization or Website is referred to in this work as a citation and/or a potential source of further information does not mean that the author or the publisher endorses the information the organization or Website may provide or recommendations it may make. Further, readers should be aware that Internet Websites listed in this work may have changed or disappeared between when this work was written and when it is read. No warranty may be created or extended by any promotional statements for this work. Neither the publisher nor the author shall be liable for any damages arising herefrom.

For general information on our other products and services or for technical support, please contact our Customer Care Department within the United States at (800) 762-2974, outside the United States at (317) 572-3993 or fax (317) 572-4002.

Wiley also publishes its books in a variety of electronic formats. Some content that appears in print may not be available in electronic formats. For more information about Wiley products, visit our web site at www.wiley.com.

Library of Congress Cataloging-in-Publication Data:

Names: Baba, Nadim Z., editor. | Guichet, David L., editor. | American College of Prosthodontists, issuing body.

Title: Journal of prosthodontics on complex restorations / edited by Nadim Z. Baba, David L. Guichet.

Other titles: Journal of Prosthodontics.

Description: Hoboken, New Jersey : John Wiley & Sons Inc. : American College of Prosthodontists, [2016] | Compilation of articles from Journal of Prosthodontics. | Includes bibliographical references and index.

Identifiers: LCCN 2016019941 | ISBN 9781119274490 (cloth) | ISBN 9781119274582 (epub) | ISBN 9781119274575 (Adobe PDF)

Subjects: | MESH: Prosthodontics—methods | Dental Restoration, Permanent—methods | Collected Works

Classification: LCC RK651 | NLM WU 500 | DDC 617.6/9—dc23 LC record available at <https://lccn.loc.gov/2016019941>

Cover image: Courtesy of Nadim Z. Baba and David L. Guichet

Cover design: Wiley

Printed in the United States of America.

10 9 8 7 6 5 4 3 2 1

CONTENTS

Preface	ix
Acknowledgments	xi

PART I MANAGEMENT OF MAXILLOFACIAL DEFECTS USING CAD/CAM TECHNOLOGY

1	Comparative Accuracy of Facial Models Fabricated Using Traditional and 3D Imaging Techniques	3
	<i>Ketu P. Lincoln, Albert Y. T. Sun, Thomas J. Prihoda, and Alan J. Sutton</i>	
2	Innovative Approach for Interim Facial Prosthesis Using Digital Technology	14
	<i>Fumi Yoshioka, Shogo Ozawa, Ikuo Hyodo, and Yoshinobu Tanaka</i>	
3	Updates on the Construction of an Eyeglass-Supported Nasal Prosthesis Using Computer-Aided Design and Rapid Prototyping Technology	20
	<i>Leonardo Ciocca, Achille Tarsitano, Claudio Marchetti, and Roberto Scotti</i>	
4	Ocular Defect Rehabilitation Using Photography and Digital Imaging: A Clinical Report	26
	<i>Muaiyed M. Buzayan, Yusnidar T. Ariffin, Norsiah Yunus, and Wan Adida Azina Binti Mahmood</i>	

PART II MANAGEMENT OF TOOTH WEAR

5	Full-Mouth Rehabilitation of a Patient with Gastroesophageal Reflux Disease: A Clinical Report	35
	<i>Juanli Guo, Glenn Reside, and Lyndon F. Cooper</i>	
6	Rehabilitation of a Bulimic Patient Using Endosteal Implants	40
	<i>Alberto Ambard and Leonard Mueninghoff</i>	

7	Fixed Prosthodontic Rehabilitation in a Wear Patient with Fabry's Disease	45
	<i>Avinash S. Bidra</i>	
8	Analyzing the Etiology of an Extremely Worn Dentition	53
	<i>Ronald G. Verrett</i>	
9	Full-Mouth Rehabilitation of a Patient with Severely Worn Dentition and Uneven Occlusal Plane: A Clinical Report	63
	<i>Elnaz Moslehifard, Sakineh Nikzad, Farideh Geraminpanah, and Farhang Mahboub</i>	
10	Implant-Supported Prosthetic Rehabilitation of a Patient with Localized Severe Attrition: A Clinical Report	73
	<i>Isil Cekic-Nagas and Gulfem Ergun</i>	
11	Restoration of the Occlusal Vertical Dimension with an Overlay Removable Partial Denture: A Clinical Report	81
	<i>Piero Rocha Zanardi, Mayara Silva Santos, Roberto Chaib Stegun, Newton Sesma, Bruno Costa, and Dalva Cruz Laganá</i>	

PART III MANAGEMENT OF CONGENITAL DISORDERS

12	Full-Mouth Rehabilitation of a Hypohidrotic Ectodermal Dysplasia Patient with Dental Implants: A Clinical Report	89
	<i>Afsaneh Shahrokhi Rad, Hakimeh Siadat, Abbas Monzavi, and Amir-ali Mangoli</i>	
13	Prosthetic Rehabilitation with Collapsible Hybrid Acrylic Resin and Permanent Silicone Soft Liner Complete Denture of a Patient with Scleroderma-Induced Microstomia	94
	<i>Kunwarjeet Singh, Nidhi Gupta, Ridhima Gupta, and Dex Abraham</i>	
14	Nasoalveolar Molding and Long-Term Postsurgical Esthetics for Unilateral Cleft Lip/Palate: 5-Year Follow-Up	100
	<i>Pravinkumar G. Patil, Smita P. Patil, and Soumil Sarin</i>	
15	Interdisciplinary Care for a Patient with Amelogenesis Imperfecta: A Clinical Report	107
	<i>Catherine Millet, Jean-Pierre Duprez, Christine Khoury, Laurent Morgon, and Béatrice Richard</i>	
16	Maxillary Rehabilitation Using a Removable Partial Denture with Attachments in a Cleft Lip and Palate Patient: A Clinical Report	116
	<i>Marina Rechden Lobato Palmeiro, Caroline Scheeren Piffer, Vivian Martins Brunetto, Paulo César Maccari, and Rosemary Sadami Arai Shinkai</i>	
17	Treatment of a Patient with Cleidocranial Dysplasia Using a Single-Stage Implant Protocol	121
	<i>Vicki C. Petropoulos, Thomas J. Balshi, Glenn J. Wolfinger, and Stephen F. Balshi</i>	
18	Full-Mouth Rehabilitation of an Edentulous Patient with Papillon-Lefèvre Syndrome Using Dental Implants: A Clinical Report	128
	<i>Leila Ahmadian, Abbas Monzavi, Rasoul Arbabi, and Hamid Mahmood Hashemi</i>	

PART IV MANAGEMENT OF ORTHODONTIC/PROSTHODONTIC PATIENTS

- 19 Aspects of Orthodontic-Prosthetic Rehabilitation of Dentofacial Anomalies** **137**
Zorica Ajdukovic, Mirjana Janosevic, Gordana Filipovic, Stojanka Arsic, Predrag Janosevic, and Nenad Petrovic
- 20 Orthodontic Treatment and Implant-Prosthetic Rehabilitation of a Partially Edentulous Patient** **148**
Milton M.B. Farret, Marcel Marchiori Farret, Jhosué Carlesso, and Oscar Carlesso
- 21 Prosthetic Rehabilitation of a Cleidocranial Dysplasia Patient with Vertical Maxillofacial Deficiency: A Clinical Report** **153**
Kwantae Noh, Kung-Rock Kwon, Hyowon Ahn, Janghyun Paek, and Ahran Pae
- 22 Occlusal Rehabilitation of Pseudo-Class III Patient** **161**
Antônio Carlos Cardoso, Cimara Fortes Ferreira, Elisa Oderich, Moira Leão Pedrosa, and Russell Wicks

PART V MANAGEMENT OF PATIENTS WITH SURGICAL AND PROSTHODONTIC CHALLENGES

- 23 Alveolar Distraction Osteogenesis of the Severely Atrophic Anterior Maxilla: Surgical and Prosthetic Challenges** **169**
Hani Braidy and Marc Appelbaum
- 24 Rehabilitation of a Completely Edentulous Patient with Nonreducible Bilateral Anterior Dislocation of the Temporomandibular Joint: A Prosthodontic Challenge—Clinical Report** **175**
Moath Momani, Mohamed-Nur Abdallah, Derar Al-Sebaie, and Faleh Tamimi
- 25 Prosthodontic Rehabilitation of a Shotgun Injury: A Patient Report** **180**
Kianoosh Torabi, Ahmad H. Ahangari, Mahrou Vojdani, and Farnaz Fattahi
- 26 Surgical and Prosthodontic Treatment of a Patient with Significant Trauma to the Middle and Lower Face Secondary to a Gunshot Wound: A Clinical Report** **186**
Paul Kelly and Carl J. Drago
- 27 Treatment of Hemi-Mandibulectomy Defect with Implant-Supported Telescopic Removable Prosthesis: A Clinical Report** **199**
Athanasios Ntounis, Michael Patras, Stavros Pelekanos, and Gregory Polyzois

PART VI MANAGEMENT OF COMPLETELY EDENTULOUS PATIENTS USING NEW CERAMIC MATERIAL

- 28 The Use of Custom-Milled Zirconia Teeth to Address Tooth Abrasion in Complete Dentures: A Clinical Report** **207**
Joanne M. Livaditis and Gus J. Livaditis

29	Retrievable Metal Ceramic Implant-Supported Fixed Prostheses with Milled Titanium Frameworks and All-Ceramic Crowns: Retrospective Clinical Study with up to 10 Years of Follow-Up	214
	<i>Paulo Malo, Miguel de Araujo Nobre, Joao Borges, and Ricardo Almeida</i>	
30	Implant-Supported Fixed Dental Prostheses with CAD/CAM-Fabricated Porcelain Crown and Zirconia-Based Framework	225
	<i>Masayuki Takaba, Shinpei Tanaka, Yuichi Ishiura, and Kazuyoshi Baba</i>	
31	Full Zirconia Fixed Detachable Implant-Retained Restorations Manufactured from Monolithic Zirconia: Clinical Report after Two Years in Service	232
	<i>Fernando Rojas-Vizcaya</i>	
32	Maxillary Full-Arch Immediately Loaded Implant-Supported Fixed Prosthesis Designed and Produced by Photogrammetry and Digital Printing: A Clinical Report	241
	<i>David Penarrocha-Oltra, Ruben Agustin-Panadero, Guillermo Pradies, Sonia Gomar-Vercher, and Miguel Penarrocha-Diago</i>	
33	Full-Arch, Implant-Supported Monolithic Zirconia Rehabilitations: Pilot Clinical Evaluation of Wear Against Natural or Composite Teeth	250
	<i>Paolo Cardelli, Francesco Pio Manobianco, Nicola Serafini, Giovanna Murmura, and Florian Beuer</i>	
34	Double Full-Arch Versus Single Full-Arch, Four Implant-Supported Rehabilitations: A Retrospective, 5-Year Cohort Study	256
	<i>Paulo Maló, Miguel De Araújo Nobre, Armando Lopes, and Rolando Rodrigues</i>	
	Index	265

PREFACE

Patients often present with a myriad of complex dental restorative challenges. Prosthodontists, restorative dentists, and other team members are called upon to solve these dilemmas. In most instances, established protocols are available. Often more imaginative techniques are applied in the manufacture and the delivery of complex restorations including nontraditional workflows and new CAD/CAM solutions. Examples of this are the advances in 3D manufacturing technology, which has come of age along with dental digital design software and machinable high-strength esthetic materials.

As new information, materials, and techniques have become available, contributors to the *Journal of Prosthodontics* have shared their developments. These authors have endeavored to position their work in the context of traditional approaches to enable the reader to draw on the historical standards and differentiate the new from the tried and true. This book is a collection of notable works on the subject of complex dental restorations.

The editors have identified six areas where significant developments have taken place in the arena of complex dental restorations. The selected articles illustrate either highly refined, traditional techniques or highlight the use of new, innovative processes. These articles represent the best application of techniques available for patients requiring complex dental restorations. The foundational principles of

successful prosthodontic rehabilitations serve as a framework against which all of these articles have been evaluated.

The six areas include management of maxillofacial defects using CAD/CAM technology; management of tooth wear; management of congenital disorders; management of orthodontic/prosthodontic patients; management of patients with surgical and prosthodontic challenges; and management of completely edentulous patients using new ceramic materials.

The American College of Prosthodontists has undertaken numerous initiatives advancing knowledge in the field of complex dental restorations including digital dental rehabilitation. These include updates at the ACP Annual Session, the Prosthodontic Review Course, and the ACP Digital Dentistry Symposium, as well as efforts to support the integration of CAD/CAM technology into dental education and create a new curriculum in digital dentistry.

These initiatives are part of a strong foundation being established by the American College of Prosthodontists as it leads this arena of dentistry into the future. Together with these efforts, this compilation serves to advance scientific knowledge in the field of fundamental prosthodontics, digital dentistry, and complex dental restorations.

NADIM Z. BABA, DMD, MSD, FACP
DAVID L. GUICHET, DDS, FACP

ACKNOWLEDGMENTS

The editors would like to thank the authors whose work is collected in this volume dedicated to the restoration of complex patient situations. We wish to express our appreciation to Dr. David Felton, editor-in-chief, and Dr. Radi Masri, associate editor-in-chief of the *Journal of Prosthodontics*, for their trust in choosing us to edit this book. We take the opportunity to thank Mr. Mark Heiden for his

professionalism, help, and guidance during the preparation of this book, as well as Ms. Alethea Gerding for her work as Managing Editor of the *Journal of Prosthodontics*. We extend our thanks to Ms. Nancy Deal Chandler and the ACP executive committee for their support in bringing this book to life.

PART I

MANAGEMENT OF MAXILLOFACIAL DEFECTS USING CAD/CAM TECHNOLOGY

COMPARATIVE ACCURACY OF FACIAL MODELS FABRICATED USING TRADITIONAL AND 3D IMAGING TECHNIQUES

KETU P. LINCOLN, DMD,¹ ALBERT Y. T. SUN, PHD, CSWA,² THOMAS J. PRIHODA, PHD,³ AND ALAN J. SUTTON, DDS, MS, FACP⁴

¹Department of Graduate Prosthodontics, USAF, Joint Base San Antonio-Lackland, San Antonio, TX

²Department of Mechanical Engineering, National Taipei University of Technology, Taipei, Taiwan

³Department of Pathology, University of Texas Health Science Center, San Antonio, TX

⁴Department of Restorative Dentistry, University of Colorado School of Dental Medicine, Aurora, CO

Keywords

Moulage; facial prosthetics; 3D imaging; 3D models; dental materials; stereolithography; rapid prototyping.

Correspondence

Alan Sutton, 13045 E. 17th Ave Ste F845, Aurora, CO 80045. E-mail: Alan.sutton@ucdenver.edu

The authors deny any conflicts of interest.

Accepted June 14, 2015

Published in *Journal of Prosthodontics* September 2015

doi: 10.1111/jopr.12358

ABSTRACT

Purpose: The purpose of this investigation was to compare the accuracy of facial models fabricated using facial moulage impression methods to the three-dimensional printed (3DP) fabrication methods using soft tissue images obtained from cone beam computed tomography (CBCT) and 3D stereophotogrammetry (3D-SPG) scans.

Materials and Methods: A reference phantom model was fabricated using a 3D-SPG image of a human control form with ten fiducial markers placed on common anthropometric landmarks. This image was converted into the investigation control phantom model (CPM) using 3DP methods. The CPM was attached to a camera tripod for ease of image capture. Three CBCT and three 3D-SPG images of the CPM were captured. The DICOM and STL files from the three 3DMD and three CBCT were imported to the 3DP, and six testing models were made. Reversible hydrocolloid and dental stone were used to make three facial moulages of the CPM, and the impressions/casts were poured in type IV gypsum dental stone. A coordinate measuring machine (CMM) was used to measure the distances between each of the ten fiducial markers. Each measurement was made using one point as a static reference to the other nine points. The same measuring procedures were accomplished on all

specimens. All measurements were compared between specimens and the control. The data were analyzed using ANOVA and Tukey pairwise comparison of the raters, methods, and fiducial markers.

Results: The ANOVA multiple comparisons showed significant difference among the three methods ($p < 0.05$). Further, the interaction of methods versus fiducial markers also showed significant difference ($p < 0.05$). The

CBCT and facial moulage method showed the greatest accuracy.

Conclusions: 3DP models fabricated using 3D-SPG showed statistical difference in comparison to the models fabricated using the traditional method of facial moulage and 3DP models fabricated from CBCT imaging. 3DP models fabricated using 3D-SPG were less accurate than the CPM and models fabricated using facial moulage and CBCT imaging techniques.

Craniofacial dysmorphism (CD) is the study of structural defects caused by trauma, treatment of neoplasms, or congenital anomalies characterized by complex irregularities in the shape and configuration of facial soft tissue structures.¹ Patients with CD may undergo extensive surgical procedures, including the fabrication of facial prostheses to restore an extraoral maxillofacial defect.² The facial prostheses are not functional, but provide the patient with an esthetic result for psychological and social acceptance.^{3–6}

Anthropometry is a way to assess changes in facial soft tissue over time through line measurements between two landmarks.⁷ The challenge has been to identify landmarks and plot them accurately in the three planes of space, in order to describe the dimensions of the face.⁸ Traditionally, direct anthropometry was done using calipers. This assessment was a reliable and inexpensive method for data collection of surface measurements.⁴ However, there were several limitations, including technician training, direct patient contact requiring extensive time to make multiple measurements, patient compliance to sit in one position, inability to archive information, difficulty attaining several measurements as tissue undergoes changes with time, and finally comparing tissue changes with accurate landmark location.⁹

Making a facial moulage impression was, and still is, another means for 3D facial structure capture, analysis, and documentation. This method has been used successfully for almost 100 years, dating back to World War I.¹⁰ Currently, various impression materials like alginate, poly(vinyl siloxane), and reversible hydrocolloid are used to create a facial moulage. The facial moulage method can be time consuming, and soft tissue deformation is a significant problem. Furthermore, it is difficult to obtain accurate impressions of certain defects involving the orbit where the periorbital tissue displaces easily.¹¹ The casts made from the impressions are fragile and require large physical storage space, and it is extremely difficult to communicate physical data to other providers in distant locations.¹² Also, archival preoperative casts may not be available for many patient treatments due to storage limitation.

Several types of 3D imaging systems have been created in the past three decades, including cone beam computed

tomography (CBCT) and 3D stereophotogrammetry (3D-SPG). Both methods are noninvasive and allow for archival of data and virtual models that can subsequently be used for comparison purposes.

Computed tomography (CT), and more specifically CBCT, is currently used to capture soft tissue surface images because it is accurate and repeatable for anthropometric measurements.⁷ Collimating the X-ray beam decreases the radiation exposure dose, and the scan time is 10 to 70 seconds.¹³ The dose of radiation ranges between 60 and 1000 μSv versus medical grade CT of the mandible, which ranges from 1320 to 3324 μSv .^{13–15} More recent studies have generated 3D facial soft tissue surface computer models from image data captured by CBCT. Linear anthropometric measurements on computer models using CBCT software proved reliable and as accurate as the traditional direct method.⁷ The data and virtual models are easily archived without physical storage requirements and can provide pre- and postoperative information for skeletal or soft tissue comparisons.¹³

3D-SPG is a newer technique/method for craniofacial surface imaging that allows for the capture evaluation of the external surface of a subject. The method creates a 3D image reconstructed from multiple digital images taken at different angles simultaneously. The resultant image is a collection of points positioned along an x , y , and z coordinate system. These points can be identified as landmarks, then used for subsequent analysis.⁹ Reports indicate that 3D-SPG is reliable and accurate for determining the location of landmarks and interlandmark craniofacial distances.^{16,17} The advantages include minimal artifact production due to short image capture time (approximately 1.5 ms), ability to archive and compare subject images, three-point (x , y , z) coordinate format of locating tissue landmarks, high resolution, and no radiation. Software programs are available to identify landmarks and calculate anthropometric measurements.¹⁸ In addition, the error in the location of a landmark when using 3D-SPG is less than 1 mm.¹⁹

The use of 3D-SPG has a great potential for use in the military. During World War II, the Korean War, and the Vietnam War, the mean incidence of head, face, and neck injury (HNFI) was approximately 16%. A recent study

looked at the characteristics and causes of HFNI sustained by US military forces during the stability and support phase of Operation Iraqi Freedom (OIF-II). The number of HFNI increased to 39%, and of these injuries, 65% were injuries to the face.²⁰ A more recent study showed a comprehensive analysis of craniomaxillofacial battle injuries sustained by military members evacuated to level III-V military treatment facilities to be 42.2% HNFIs.²¹ The reason for the notable increase in the past decade is an increase in survival rate due to improvement in body armor, battlefield medicine, tactically placed medical units, and quick evacuation tactics.

Both CBCT and 3D-SPG use computer-aided design (CAD) software to facilitate the design of soft tissue surface images and virtual models. With rapid prototyping (RP), information from the CAD software can be used along with computer-aided manufacturing (CAM) to fabricate 3D physical models. Image data from CBCT and 3D-SPG scans translated into the digital imaging and communication in medicine (DICOM) file format, which are converted to a CAM file format to produce a 3D model using RP methods and equipment.

One RP method used in the medical and dental field is 3D printing (3DP). This process uses a polyjet selectively depositing fine powder polymer droplets evenly along a piston and liquid binder. Additional layers are added as the piston powder bed and cured model is lowered layer by layer. The resolution accuracy is 100 μm for one-dimensional features and 300 μm for 3D features.¹¹ The 3D printed models are accurate to 0.016 mm (Objet Eden 260V; Stratasys Ltd., Minneapolis, MN), and the build time is at a rate of 1 cm of height per hour.²²

The 3D models are useful for surgical planning, creation of surgical templates, and fabrication of craniofacial prostheses and custom implants used in craniofacial reconstruction.

The accuracy of the RP models has been measured by software calculations,^{23–25} digital calipers,¹⁸ and more recently the use of a coordinate measurement machine (CMM). The CMM can provide accurate location of x , y , and z coordinate reference points. This device is very useful in locating the same landmark on various models and therefore accurate in determining any error in model production.

RP techniques are proving beneficial in the treatment planning, diagnosis, surgical assistance, prosthesis fabrication and postassessment of patients with craniofacial anomalies, facial trauma, and structural defects caused by neoplasms; however, further studies need to be done to evaluate the precise fit of models fabricated from soft tissue imaging. The purpose of this investigation was to compare the accuracy of facial models fabricated using facial moulage impression methods to the 3DP fabrication methods using soft tissue images obtained from CBCT and 3D-SPG.

MATERIALS AND METHODS

One human form was obtained from a two-pod 3D-SPG surface imaging system and software system (3dMDface; 3dMD, Atlanta, GA). The scanned image was saved as an Standard Triangulation Language (STL) file and uploaded into the modeling software program (Geomagic Freeform Modeling Plus; Geomagic, Wilmington, MA) to create the virtual model. The virtual model was used to design the control phantom model (CPM). Five millimeter diameter spheres were built into the model to mark the following ten landmarks on the facial soft tissue: Glabella, Nasion, Pronasale, right and left Orbitale, right and left Frontale, right and left Cheilion, and Pogonion (Fig 1.1A).^{4,5,7,9,26,27} The




FIGURE 1.1 (A) Virtual model with landmarks. (B) Printed control phantom model (CPM) frontal view. (C) CPM lateral view.

virtual master model was processed using 3DP software (Objet Studio; Stratasys Ltd., Minneapolis, MN), and the physical CPM was created using 3DP (Objet Eden 260V; Stratasys Ltd.) (Figs 1.1B and 1.1C).

The CPM was used to create three facial moulage experimental gypsum dental stone models using reversible hydrocolloid (Polyflex Duplicating Material; Dentsply International, York, PA). Reversible hydrocolloid at room temperature was heated to its liquefaction temperature to convert the gel to the sol condition.²⁸ The reversible hydrocolloid was applied to the CPM using a synthetic brush (Synthetic brush #16; Dentsply). Cotton gauze (2 in²) was embedded in the solidifying reversible hydrocolloid to reinforce the material and allow for the attachment to dental stone. A thin consistency of dental stone (Mounting Stone ISO type 3; Whip Mix Corp., Louisville, KY) was applied over the reversible hydrocolloid and gauze in a uniform half-inch thickness to fabricate an external tray. The ratio of the dental stone to filtered water was 900 g to 170 ml. Once the stone set, the impression was removed from the master model and poured in type IV dental stone (Silky Rock, Whip Mix Corp.) (Fig 1.2A).²⁹ The ratio of type IV dental stone to filtered water was 600 g to 138 ml. Each of the resultant three stone models was labeled accordingly (Fig 1.2B).

To position the CPM for CBCT capture, a tripod measurement base assembly was fabricated using a tripod screw platform with acrylic resin (Ortho Acrylic Resin; Great Lakes Orthodontics, Tonawanda, NY) (Fig 1.3).³⁰ The CBCT system (Kodak 9500 Cone Beam 3D System; Carestream Health, Inc., Rochester, NY) was calibrated following the manufacturer's instructions. The CPM was stabilized on the tripod, and a total of three images were made individually and labeled one through three (Fig 1.4). The images were saved as DICOM files and copied onto a disc for use with the RP system.

The same tripod base assembly for the CPM was used to obtain the 3D-SPG images (3dMDface). The system was calibrated according to the manufacturer's instructions.³¹ The tripod with CPM was positioned at a 15° anterior tilt to capture an image with minimal shadowing (Fig 1.5). A total of three images were made individually and labeled one through three. These images were saved as STL files and saved onto a disc for use with the RP system.

The DICOM images and STL files from the CBCT and 3dMD, respectively, were used to create the virtual models using computer software. A DICOM segmentation program (MIMICS 12.1; Materialise Dental, Plymouth, MI) was used to identify the CPM and generate a surface model (in STL format) from the series of CBCT images (Fig 1.6). The six STL files were aligned, and a common base was designed and merged to the 3D surface of each of the scans. Then, using 3DP (Objet Eden 260V) six individual models were fabricated prior to measurement procedures (Fig 1.7).




FIGURE 1.2 (A) Reversible hydrocolloid impression. (B) Stone model.

The printed RP models and gypsum stone models were measured for accuracy by three individual raters using a CMM (Faro, Lake Mary, FL). A 3 mm ball probe stylus was placed on the surface of each fiducial marker, and a discrete point cloud was recorded into Geomagic Studio as the measuring software interface. Each cloud data set was interpreted as a sphere feature on the model. The scans resulted in a collection of point cloud data representing the feature location in space relative to each of the other spheres (Figs 1.7 and 1.8).

The software was then used to calculate a best-fit sphere for each of the point cloud groups. Three sphere centers (1, 2, and 10) were used to define a reference plane. New points were defined by projecting each of the sphere centers to the reference plane in a direction normal to the plane (Fig 1.9).



FIGURE 1.3 Tripod assembly.

Nine projected points were then measured against the projected point #3 on the same model. The following 3D Distance Formula was used where $i=1, 2, 4, \dots, 10$ and $j=3$.

$$\sqrt{(x_i - x_j)^2 + (y_i - y_j)^2 + (z_i - z_j)^2}$$

This procedure was accomplished for each landmark on the CPM model and compared to the same measurements obtained on the facial moulage stone reproductions and the printed models.

Statistical Analysis

Vertical distances from point #3 were analyzed. First, the master distances were averaged over the three raters at each point. Then, for each rater on CBCT (CT), 3D-SPG (OP), and stone (ST) the vertical distance from point #3 was subtracted from the master mean of the three raters and divided by the master mean and multiplied by 100 to obtain a percent difference from the master for each rater and each point of the other three methods. The percent differences were



FIGURE 1.4 Tripod positioning for CBCT capture.

analyzed with ANOVA for repeated measures, since the raters repeatedly measured each point from each of the cast methods.

The three raters were compared, the three methods were compared, and the nine points were compared. Since all distances were relative to point #3, those values were always zero, and that point was not included in the analysis. In addition, the two-way interaction of method and points was analyzed. Interactions with the rater factor were the denominators for the F-tests in the ANOVA for repeated measures. Tukey's comparison was done for pairwise comparisons of means following the ANOVA. Residuals of the ANOVA were calculated and plotted to verify they had a near normal bell-shaped curve and that their variance was similar over the range of predicted values.

RESULTS

The percent difference of each of the three methods (CT, OP, ST) from the control mean relative to point 3 for each method is shown in Figure 1.10. The ANOVA was done and is displayed in Table 1.1. The method (meth) by point (pt)




FIGURE 1.5 (A) Tripod positioning of CPM for 3dMD capture. (B) 15° anterior tilt position of CPM.

source was significant. Therefore, a comparison was done with the three methods to the control and each method to each other. Multiple comparisons showed the raters were not different. Pairwise comparisons of the methods were different with OP not having as small a percentage error as the other two methods, while the other two had similar percent error overall (Tables 1.2 and 1.3). Overall, OP showed statistically significant difference ($p < 0.05$) in comparison to the CT and ST; however, Figure 1.10 shows the greatest difference localized to points #1, 2, and 5. The OP data for the other points are similar to the CT and ST findings.

DISCUSSION

The impression technique for making facial models has been used for many years, but a major disadvantage is soft tissue deformation caused by the direct contact of the impression material to the facial soft tissue. Holberg et al reported that making an alginate impression of the face produced between 1 and 3 mm of soft tissue deformation in varying areas.³² Germec-Cakan et al found significant differences between clinical and facial plaster cast measurements explained by distortion related to the impression material.¹² In this study, the CPM was made from a rigid resin material. When a facial moulage was made of the CPM, there were no signs of deformation, which would normally be seen in a patient. Therefore, the results at each point showed very minimal percentage difference in comparison to the CPM.




FIGURE 1.6 Mimics DICOM segmentation.




FIGURE 1.7 Point cloud data of spheres 1 to 4.

Numerous studies have shown that imaging done by CT, and more specifically CBCT, is currently used to capture hard and soft tissue surface images because of its accuracy, reliability, and repeatability for anthropometric measurements.^{33–35} Fourie et al compared linear measurements derived from 11 soft tissue landmarks on seven cadaver heads made directly using digital calipers to CBCT-based computer-generated models. Their results showed surface detail of the soft tissue images was insufficient; however, overall, the data proved to be reliable and accurate.⁷ Once again, the CPM was a rigid resin form without soft tissue-like surfaces. The results showed that CT data were not significantly different from the CPM measurements and confirmed that CT-generated models were reliable and accurate.




FIGURE 1.8 Point cloud data of spheres 1 to 10.




FIGURE 1.9 Defined reference plane using sphere centers of 1, 2, and 10.

3D-SPG imaging provides accurate detail of surface texture, contour, and color. Many studies have been done comparing 3D-SPG imaging to direct anthropometry resulting in repeatable, precise, and accurate measurements.^{16,36,37} Additional studies by Weinberg et al and Wong et al showed an increase in accuracy and precision of landmark location with labeling prior to image capture.^{5,9} In a study done by Plooi et al, midline landmarks were precisely generated compared to pair landmarks, especially if the interlandmark distance increased.¹⁷ In the present study, spherical landmarks were created in the CPM that were reproducible using impression material and the two imaging techniques. Capturing the landmarks was an important factor to calculate the linear measurements and make comparisons of accuracy.

In all of these studies involving CT or 3D-SPG computer-generated images, 3D imaging software was used to calculate linear measurements, which were compared to direct anthropometric measurements. Caliper measurements can be subjective and therefore the accuracy and reliability of the data may be questionable.³⁸ In the present study, a CMM was used to decrease the subjectivity found in using digital calipers. In Taft et al's study,³⁹ stainless steel spheres (5.00 ± 0.005 mm in diameter) were secured on a dry cadaver skull in seven locations. Point locations of the spheres were measured by placing the CMM ball probe on the points of interest, thus improving accuracy and reliability. Also, because the spheres were used as fiducial markers, a mean centroid location was identified for each sphere, and the distance between two points was then determined. In this study, a CMM in conjunction with computer software was used to calculate a best-fit sphere for each of the landmarks. There were three raters who collected point cloud data at each sphere, and a rater pairwise comparison showed no significant difference among the raters. The CMM proved to be reliable and accurate in this study.

Both CBCT and 3D-SPG image files can be imported into CAD/CAM software to create an RP model. The accuracy of the RP models have been studied by measuring distances




FIGURE 1.10 Percent difference from control mean relative to point 3.

between landmarks on 3D phantom models and comparing those points with measurements calculated through the CT software.^{23–25} In this study 3DP models were fabricated using the DICOM images and STL files from the CBCT and 3dMD, respectively; however, instead of using computer

software to make the measurements for the CBCT and 3DMD, the CMM was the constant measuring tool for the stone and the 3DP models.

The accuracy of RP models fabricated to replace hard tissue has been studied with skull models and is used in

TABLE 1.1 ANOVA Table for Percent Difference from Master Mean

Source	DF	Sum of Squares	Mean Square	<i>F</i> value	<i>Pr</i> > <i>F</i>
Rater	2	0.53050924	0.26525462	1.60	0.2169
Meth	2	5.03638366	2.51819183	59.86	0.0010
Meth × rater	4	0.16827539	0.04206885	0.25	0.9048
Pt	8	7.17435054	0.89679382	5.05	0.0029
Pt × rater	16	2.83916780	0.17744799	1.07	0.4169
Meth × pt	16	13.47849934	0.84240621	5.09	< 0.0001
Residual	32	5.29215010	0.165379691		
Corrected total	80	34.51933608			

TABLE 1.2 Rater Pairwise Comparison: Tukey Comparison

Rater	Difference Percentage Mean	<i>p</i> -Values	
		versus 2	versus 3
1	0.61600732	0.1984	0.4865
2	0.81106176		0.8187
3	0.74416230		

Table 1.3 Method Pairwise Comparison: Tukey Comparison

Method	Difference Percentage Mean	<i>p</i> -Values	
		versus OP	versus ST
CT	0.62195799	0.0030	0.1357
OP	1.06703386		0.0010
ST	0.48223953		

reconstructive craniofacial surgeries.^{33,39} However, many craniofacial anomalies also involve the facial soft tissue, requiring accurate dimensional measurements to fabricate prosthetics. Studies have been done using CT scans to create casts from the RP models for facial prosthesis engineering.^{40–43} Few studies have evaluated the precise fit of prostheses fabricated from RP models manufactured from soft tissue imaging obtained by 3D-SPG.¹⁸ This investigation was limited because a rigid resin model was used as a control.

According to this investigation, 3DP models fabricated using 3D-SPG showed statistical difference compared to the models fabricated using the traditional method of facial moulage and 3DP models fabricated from CBCT imaging. Major discrepancies stemmed from points #1 and 2 (Fig 1.10). The greater difference of the optical scan point values in comparison to those of the models made from CBCT imaging and facial moulage at the other points may be because a two-pod system was used and the 15° anterior angulation of the control during 3dMD image capture.

Percentage differences of the individual points on all three methods in comparison to the control, points #4, 8, 9, and 10 were not statistically significant differences from the CBCT and stone method. Thus, 3D-SPG is a viable option for RP production of facial models, especially in situations where it is not feasible to use CBCT imaging. Many patients with craniofacial dysmorphologies undergo numerous surgeries in a short period of time, and the 3D-SPG method would eliminate radiation exposure from CT. Additionally, the short image capture time would be extremely beneficial for patients with the inability to be still for the time it takes to make a CBCT. Comparative growth studies could be accomplished using this technology, and if necessary, RP models could be engineered to help with treatment needs. The incorporation of this technology is beneficial for the facial reconstruction process because of its high efficiency, the ability to provide accurate facial surface detail, and the overall treatment planning information obtained for patients. The ability to archive images further helps with the treatment process and analysis of any subsequent changes in the soft tissue.

The 3D-SPD method can also be used in conjunction with CBCT.⁴⁴ The accurate hard tissue image obtained from a CBCT can be referenced to a 3D-SPG scan providing detailed images relating the hard tissue with soft tissue for analysis prior to orthognathic surgery. Furthermore, it is difficult to capture an image using 3D-SPG in a defect area where undercuts are present, and the image captured through a CBCT may help define the boundaries of the defect. This merging of hard and soft tissue images can be extremely beneficial in viewing, treatment planning, and fabricating an accurate prosthesis for a craniofacial defect.

In addition, military members suffering from HFNI present to medical and dental clinics with facial dysmorphologies, such as missing ears, requiring facial prostheses. In the past, these patients required creation of models of the

area of deformity by using previous 2D photographs, an impression of family member anatomical replicas, or a prosthesis fabricated by an anaplastologist to replicate the lost tissue. Now, with 3D-SPG and CBCT images, recreation of missing tissue can be accomplished by banked images, images of family members, or even custom-created anatomic forms. Furthermore, images of military members could be obtained and archived prior to entering a military conflict. If the military member should sustain any HFNI, then the archive image can be referenced to create a model in the fabrication of a more accurate facial prosthesis.

Future Directions

There were limitations to this study. First, the CPM was made from a rigid resin material. When a facial moulage was made of the CPM, there were no signs of deformation, which would normally be seen in a patient. Also, a CT image does capture hard tissue detail accurately but lacks in soft tissue detail. 3D-SPG imaging provides a 3D viewing ability to see soft tissue color and texture detail. Therefore, future studies should be done using a patient with a craniofacial defect, and all three methods should be reinvestigated. Also, investigation of five-pod 3D-SPG systems may provide a more accurate 3D image. Finally, it may be beneficial to investigate the accuracy of 3D-SPG in conjunction with CBCT imaging to fabricate facial models.

CONCLUSION

This investigation was based on an innovative research setting creating facial models using a two-pod 3D-SPG imaging system, and a CBCT imaging method, then comparing the accuracy of these models to the traditional facial moulage impression model fabrication technique.

Within the limitations of this investigation, the following conclusions could be made:

1. 3DP models fabricated using 3D-SPG showed statistical difference in comparison to the models fabricated using the traditional method of facial moulage and 3DP models fabricated from CBCT imaging.
2. 3DP models fabricated using 3D-SPG were less accurate in comparison to the CPM and models fabricated using facial moulage and CBCT imaging techniques.
3. Models fabricated using CBCT imaging and facial moulage showed no statistical difference and proved to be accurate in comparison to the CPM.

ACKNOWLEDGMENTS

A special thanks to Mrs. Nancy Hansen, CDT, CCA, BS, and Mr. Alain Carballeyra, BS, for their support.

REFERENCES

- Merriam-Webster Dictionary. [Internet]. Merriam-Webster, Incorporated; 2013. Dysmorphology; Accurate; Precision. [cited 2013 Feb 10]. Available from: <http://www.merriam-webster.com>.
- Eckert SE, Desjardins RP: Prosthetic considerations. In Brankmark PI (ed): *Osseointegration in Craniofacial Reconstruction*, Vol 1 (ed 1). Chicago, Quintessence, 1998, pp 87–88.
- Arridge SR, Moss JP, Linney AD, et al: Three-dimensional digitization of the face and skull. *J Maxillofac Surg* 1985;13:136–143.
- Posnick JC, Farkas LG: The application of anthropometric surface measurements in craniomaxillofacial surgery. In Farkas LG (ed): *Anthropometry of the Head and Face* (ed 1). New York, Raven Press, 1994; pp 125–138.
- Weinberg SM, Scott NM, Neiswanger K, et al: Digital three-dimensional photogrammetry: evaluation of anthropometric precision and accuracy using a Genex 3D camera system. *Cleft Palate Craniofac J* 2004;41:507–518.
- Davis BK: The role of technology in facial prosthetics. *Curr Opin Otolaryngol Head Neck Surg* 2010;18:332–340.
- Fourie Z, Damstra J, Gerrits PO, et al: Accuracy and repeatability of anthropometric facial measurements using cone beam computed tomography. *Cleft Palate Craniofac J* 2011;48:623–630.
- Souccar NM, Kau CH: Methods of measuring the three-dimensional face. *Sem Orthod* 2012;18:187–192.
- Wong JY, Oh AK, Ohta E, et al: Validity and reliability of 3D craniofacial anthropometric measurements. *Cleft Palate Craniofac J* 2008;45:232–239.
- Alexander C: Faces of war: amid the horrors of World War I, a corps of artists brought hope to soldiers disfigured in the trenches. *Smithsonian Magazine* 2007;37:72.
- Beumer J, Wolfaardt J, Lee M: *Maxillofacial Rehabilitation* (ed 3). Hanover Park, IL, Quintessence, 2011, pp. 282–3, 361–3, 391–394.
- Germec-Cakan D, Canter HI, Nur B, et al: Comparison of facial soft tissue measurements on three-dimensional images and models obtained with different methods. *J Craniofac Surg* 2010;21:1393–1399.
- Scarfe WC, Farman AG, Sukovic P: Clinical applications of cone-beam computed tomography in dental practice. *J Can Dent Assoc* 2006;72:75–80.
- Kau CH, Olim S, Nguyen JT: The future of orthodontic diagnostic records. *Seminar in Orthodontics* 2011;17:39–45.
- DeVos W, Casselman J, Swennen GR: Cone-beam computerized tomography (CBCT) imaging of the oral and maxillofacial region: a systemic review of the literature. *Int J Oral Maxillofac Surg* 2009;38:609–625.
- Heike CL, Cunningham ML, Hing AV, et al: Picture perfect? Reliability of craniofacial anthropometry using three-dimensional digital stereophotogrammetry. *Plast Reconstr Surg* 2009;124:1261–1272.
- Plooji JM, Swennen GR, Rangel FA, et al: Evaluation of reproducibility and reliability of 3D soft tissue analysis using 3D stereophotogrammetry. *Int J Oral Maxillofac Surg* 2009;38:267–273.
- Sabol JV, Grant GT, Liacouras P, et al: Digital image capture and rapid prototyping of the maxillofacial defect. *J Prosthodont* 2011;20:310–314.
- Aldridge K, Boyadjiev SA, Capone GT, et al: Precision and error of three-dimensional phenotypic measures acquired from 3dMD photographic images. *Am J Med Genet A* 2005;138A:247–253.
- Wade AL, Dye JL, Mohrle CR, et al: Head, face, and, neck injuries during Operation Iraqi Freedom II: results from the US Navy-Marine Corps combat trauma registry. *J Trauma* 2007;63:836–840.
- Chan RK, Siller-Jackson A, Verrett AJ, et al: Ten years of war: a characterization of craniomaxillofacial injuries incurred during operations Enduring Freedom and Iraqi Freedom. *J Trauma* 2012;73:453–458.
- Cohen A, Laviv A, Berman P, et al: Mandibular reconstruction using stereolithographic 3-dimensional printing modeling technology. *Oral Surg Oral Med Oral Path Oral Rad Endo* Nov 2009;105:661–666.
- Fruhwald J, Schicho K, Figl M, et al: Accuracy of craniofacial measurements: computed tomography and three-dimensional computed tomography compared with stereolithographic models. *J Cranio Surg* 2008;19:22–26.
- Taylor JS: Influence of computerized tomography parameters on the quality of stereolithographic models. Biomedical Science [thesis]. San Antonio (TX): University of Texas; 1999.
- Bouma L: The accuracy of non-contact three-dimensional laser surface scanning in the fabrication of facial moulage. Biomedical Science [thesis]. San Antonio (TX): University of Texas; 2003.
- Jacobson A: *Radiographic Cephalometry; Basics of Video-imaging*. Chicago, Quintessence, 1995.
- Williams PL, Warwick R, Dyson M, et al: *Gray's Anatomy* (ed 37). New York, Churchill Livingstone, 1989.
- Shen, C: Impression materials. In Anusavice KJ (ed): *Phillip's Science of Dental Materials* (edn 11). St. Louis, Saunders, 2003, pp 231–248.
- Clarke CD: *Prosthetics*. Butler, MD, Standard Arts Press, 1965, Ch 3.
- Khambay B, Naim N, Bell A, et al: Validation and reproducibility of a high resolution three-dimensional facial imaging system. *Br J Oral Maxillofac Surg* 2008;40:27–32.
- Weinberg SM, Naidoo S, Govier DP, et al: Anthropometric precision and accuracy of digital three-dimensional photogrammetry: comparing the genex and 3dMD imaging systems with one another and with direct anthropometry. *J Craniofac* 2006;17:477–483.
- Holberg C, Schwenzer K, Mahaini L, et al: Accuracy of facial plaster casts. *Angle Orthod* 2006;76:605–611.
- Damstra J, Fourie Z, Huddleston Slater JJR, et al: Accuracy of linear measurements from cone beam computed tomography derived surface models of different voxel sizes. *Am J Orthod Dentofacial Orthop* 2010;137:16.e1–16.e6.

34. Stratemann SA, Huang JC, Maki K, et al: Comparison of cone beam computed tomography imaging with physical measures. *Dentomaxillofac Radiol* 2008;37:80–93.
35. Mischkowski R, Reinhard P, Lutz R, et al: Geometric accuracy of a newly developed cone beam device for maxillofacial imaging. *Oral Surg Oral Med Oral Path Oral Rad Endo* 2007;104:551–559.
36. Ghoddousi H, Edler R, Haers P, et al: Comparison of three methods of facial measurement. *Int J Oral Maxillofac Surg* 2007;36:250–258.
37. Lubbers HT, Medinger L, Kruse A, et al: Precision and accuracy of the 3dMD photogrammetric system in craniomaxillofacial application. *J Craniofac Surg* 2010;21:763–767.
38. Jamali AA, Deuel C, Perreira A, et al: Linear and angular measurements of computer generated models: are they accurate, valid, and reliable? *Comput Aided Surg* 2007;12:278–285.
39. Taft RM, Kondor S, Grant GT: Accuracy of Rapid Prototype models for head and neck reconstruction. *J Prosthet Dent* 2011;106:399–408.
40. Esses SJ, Berman P, Bloom A, et al: Clinical applications of physical 3D models derived from MDCT data and created by rapid prototyping. *AJR Am J Roentgenol* 2011;196:683–688.
41. Karayazgan-Saracoglu B, Gunay Y, Atay A: Fabrication of an auricular prosthesis using computed tomography and rapid prototyping technique. *J Craniofac Surg* 2009;20:1169–1172.
42. Turgut G, Sacak, KK, et al: Use of rapid prototyping in prosthetic auricular restoration. *J Craniofac Surg* 2009;20:321–325.
43. Marafon PG, Mattos BS, Saboia AC, et al: Dimensional accuracy of computer-aided design/computer-assisted manufactured orbital prostheses. *Int J Prosthodont* 2010;23:271–276.
44. Maal TJ, Plooij JM, Rangel FA, et al: The accuracy of matching three-dimensional photographs with skin surfaces derived from cone-beam computed tomography. *Int J Oral Maxillofac Surg* 2008;37:641–646.

INNOVATIVE APPROACH FOR INTERIM FACIAL PROSTHESIS USING DIGITAL TECHNOLOGY

FUMI YOSHIOKA, DDS, PHD,¹ SHOGO OZAWA, DDS, PHD,¹ IKUO HYODO, MD,² AND YOSHINOBU TANAKA, DDS, PHD¹

¹Department of Removable Prosthodontics, School of Dentistry, Aichi Gakuin University, Nagoya, Japan

²Department of Plastic and Reconstructive Surgery, Aichi Cancer Center, Nagoya, Japan

Keywords

Facial prosthesis; interim prosthesis; rapid prototyping; magnetic attachment; surgical prosthesis.

Correspondence

Fumi Yoshioka, 2-11 Suemuridori, Chikusa-ku, Nagoya, Aichi, 4648651, Japan. E-mail: fumi@dpc.agu.ac.jp

Supported by the Hori Sciences and Arts Foundation.

Presented at the Advanced Digital Technology in Head and Neck Reconstruction, 4th International Conference, May 2012; Freiburg, Germany.

The authors deny any conflicts of interest.

Accepted February 20, 2015

Published in *Journal of Prosthodontics* August 2015

doi: 10.1111/jopr.12338

ABSTRACT

Despite the important role of facial prosthetic treatment in the rehabilitation of head and neck cancer patients, delay in its implementation can be unavoidable, preventing patients from receiving a prompt facial prosthesis and resuming a normal social life. Here, we introduce an innovative method for the fabrication of an interim facial prosthesis. Using a 3D modeling system, we simplified the fabrication method and used a titanium reconstruction plate for facial prosthesis retention. The patient received the facial prosthesis immediately after surgery and resumed a normal social life earlier than is typically observed with conventional facial prosthetic treatment.

A facial prosthesis is a treatment option for repairing facial defects in head and neck cancer patients. Following surgery, facial prosthesis fabrication is often delayed due to a number

of factors. For one, patients are not referred to a prosthodontist until the surgeon has confirmed that the surgical site has completely healed with no tumor recurrence. In addition,

prosthodontists do not initiate treatment until the wound has completely healed, as modification of silicone material is difficult once it has cured; however, physicians should carefully consider the psychological impact following loss of part of the face, even in the wound-healing period.¹

Facial prosthesis fabrication via the conventional method consists of many steps, as follows: making an impression, building a wax sculpture for the missing structure, converting it into silicone material, and matching intrinsic and extrinsic coloration with the skin. To reduce the time required to construct a prosthesis, a rapid prototyping (RP) system has been applied to prosthodontic treatment² and facial prosthesis fabrication.^{3–7} We applied a 3D digitizer and RP system to the fabrication of orbital prostheses⁷ and found that treatment time was markedly reduced.

This new, rapid system has since been used for the simple fabrication of facial prostheses,⁸ supporting the potential to deliver facial prostheses immediately after surgery to repair the facial defect and decrease the psychological depression of patients. Here, we introduce an innovative method for the immediate fabrication of facial prostheses using a 3D RP system.

CLINICAL REPORT

A 44-year-old woman with nasal cancer was scheduled to undergo a rhinectomy. The histopathologic nature was squamous cell carcinoma. Induction or adjuvant chemotherapy or radiotherapy was not planned. On surgical planning, however, nasal reconstructive surgery was deemed insufficient by the plastic surgeon. The patient was therefore referred to a maxillofacial prosthodontist as a candidate for prosthodontic

treatment in combination with lip reconstruction surgery. The patient was informed that the resected area would be extended to the external nose, upper lip, and bilateral maxilla on presurgical planning. A combination of myocutaneous forearm free-flap reconstruction and facial prosthesis of the upper lip and nose was planned for the midfacial defect. A titanium reconstruction plate (UCMF primary reconstruction plate; Stryker Corp., Kalamazoo, MI) was designed for the retention of an interim nasal prosthesis.

Presurgical Consult and Data Acquisition

Before surgery, a digital image was acquired using a 3D digitizer (Rexcan 3; Solutionix Co., Seoul, South Korea). This 3D digitizer consisted of an industrial 3D scanner using phase-shifting optical triangulation and CCD twin-camera technology, and enabled 3D data to be acquired within a short duration (2 seconds) and purported excellent accuracy (± 0.001 mm). Computed tomography (CT) data for the diagnosis were used for data establishment. Both sets of digital data were converted into STL format and aligned using 3D transforming software (Geomagic Studio; Geomagic, Morrisville, NC; Fig 2.1). The surgeon then estimated the ablation site by referencing CT data, prosthodontists designed a virtual prosthesis with reference to scanned data, and CT data were aligned using CAD software (Free-Form Modeling; Sensable, Wilmington, MA; Fig 2.2). A facial model combined with the designed prosthesis was manufactured with dental stone (New Plastone 2; GC Corp., Tokyo, Japan) using a 3D milling machine (MDX-40; Roland DG, Shizuoka-ken, Japan). A surgical template made from a formed thermoplastic polyethylene terephthalate glycol plate (Erkodur; Erkodent, Pfalzgrafenweiler,




FIGURE 2.1 Schema of 3D image establishment. Scanned surface data and CT data were aligned.




FIGURE 2.2 Estimated resected site and designed virtual facial prosthesis in place (perforated line).

Germany) was fabricated using a vacuum-forming machine (Erkopress; Erkodent; Fig 2.3).

During surgical reconstruction, the titanium reconstruction plate was bent and inserted at an appropriate position, where the retentive force of the prosthesis is most effective, and the plate would not disturb the morphology of the prosthesis. The surgical template was used to determine the position by referring to presurgical planning (Fig 2.4), and the plate was fixed on the anterior wall of the maxilla. The template was also used as reference to position the forearm flap.

Postsurgical Data Acquisition and Fabrication of Facial Prosthesis

Following surgery, another digital image was acquired using the 3D digitizer, and a postoperative facial mold was manufactured using the 3D milling machine (Fig 2.5). A post-operative template made of formed thermoplastic polymethylmethacrylate (PMMA) plate was fabricated using the vacuum-forming machine. Pre- and postsurgical templates were combined to make the substructure of the prosthesis. The outer side of the substructure was then covered

with a thin layer of wax to represent skin color and texture (Fig 2.6). Skin-colored wax was used for sculpting, which is better and easier for us and the patient to capture the form of the prosthesis during sculpting when tried in.

The prototype was invested, and a silicone adhesive was applied to the surface of the substructure. The wax surface was converted into intrinsically colored silicone material (A2186F; Factor II, Lakeside, AZ). Following polymerization, the facial prosthesis was removed from the mold and tried on the patient. Magnetic keepers (Gigauss D800; GC Corp.) were attached to the titanium reconstruction plate (Fig 2.7), and magnetic assemblies were fixed onto the inner side of the prosthesis using self-cured resin. In this case, it took a week for initial wound healing when the resected area had been covered with the gauze. One week after surgery, an interim facial prosthesis was delivered to the patient, and the fit and orientation to the adjacent tissue as well as the retention were assessed (Fig 2.8). The side of the prosthesis facing the tissue was made of a thin plastic plate to enable modification using self-curing resin to adapt to wound healing. The outer side of the prosthesis was constructed from silicone material to represent skin color and texture. The patient was instructed to clean the plate with swabs,




FIGURE 2.3 Planned and manufactured models (left and center) and fabricated surgical template.